

Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ D-Glucose

REF	Art. No.:	E1210
LOT	Lot:	200747
	Expiry:	2021-02

Test	Specifications	Results
Part 1: Specific activity	Hexokinase ≥ 290 U/ml G6-P-DH ≥ 145 U/ml	corresponds corresponds
Part 2: pH value	Bottle #1 7.4 - 7.8	corresponds
Part 3: Solubilisation time	Bottle #1 ≤ 3 min	corresponds
Performance test	Target values $\pm 5\%$	corresponds

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2019-07-23



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com