


Quality Assurance Certificate

r-biopharm®



Product: Enzytec™ D-Gluconic Acid

REF	Art. No.:	E1223
LOT	Lot:	200793
	Expiry:	2022-12

Test	Specifications	Results
Part 1: Specific activity	6-P-Gluconate-DH \geq 220 U/ml Gluconate kinase \geq 26 U/ml	corresponds corresponds
Part 2: pH value	Bottle #1 7.4 - 7.8	corresponds
Part 3: Solubilisation time	Bottle #1 \leq 3 min	corresponds
Performance test	Target values \pm 5%	corresponds

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-05-26

Quality
Management
System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com