

# Quality Assurance Certificate

r-biopharm®



**Product:** RIDA® CUBE L-Lactic acid

<b>REF</b>	Art. No.:	RCS4260
<b>LOT</b>	Lot:	2011-1
	Expiry:	2022-02

Content	Lot	Expiry
Reagent 1	2011-1	2022-02
Reagent 2	2011-1	2022-02
RFID Card	2011-1	2022-02
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	900	927	103.0
Sample 2	140	132	94.3
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	140	134	95.7
Sample 2	35	34	97.1
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-12-03



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com