

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Acetic Acid

REF	Art. No.:	RCS4226
LOT	Lot:	2005-1
	Expiry:	2021-07

Content	Lot	Expiry
Reagent 1	2005-1	2021-07
Reagent 2	2005-1	2021-07
Reagent 3	2005-1	2021-07
RFID Card	2005-1	2021-07
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	525.0	508	96.8
Sample 2	143.0	127	88.8
Reagent blank (water sample)	0.0	1	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	100.1	100	99.9
Sample 2	28.6	25	87.4
Reagent blank (water sample)	0.0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-06-02

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality Management System certified
• ISO 9001
• ISO 13485