

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE L-Malic Acid**

<b>REF</b>	Art. No.:	RCS4280
<b>LOT</b>	Lot:	1910-3
	Expiry:	2020-11

Content	Lot	Expiry
Reagent 1	1910-3	2020-11
Reagent 2	1910-3	2020-11
RFID Card	1910-3	2020-11
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1150.0	1120.0	97.4
Sample 2	147.0	145.0	98.6
Reagent blank (water sample)	0	20	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	230.0	226.0	98.3
Sample 2	36.8	37.0	100.5
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2019-11-05

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

Quality Management System certified  
• ISO 9001  
• ISO 13485