

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE L-Malic acid

REF	Art. No.:	RCS4280
LOT	Lot:	2010-4
	Expiry:	2022-02

Content	Lot	Expiry
Reagent 1	2010-4	2022-02
Reagent 2	2010-4	2022-02
RFID Card	2010-4	2022-02
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1150.0	1134.0	98.6
Sample 2	147.0	150.0	102.0
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	230.0	215.0	93.5
Sample 2	36.8	34.0	92.4
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-11-09

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com

