

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE D-Glucose**

<b>REF</b>	Art. No.:	RCS4140
<b>LOT</b>	Lot:	2104-4
	Expiry:	2022-06

Content	Lot	Expiry
Reagent 1	2104-4	2022-06
Reagent 2	2104-4	2022-06
RFID Card	2104-4	2022-06
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1666.7	1576	94.6
Sample 2	250.0	247	98.8
Reagent blank (water sample)	0	1	passed

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	400.0	381	95.3
Sample 2	50.0	49	98.0
Reagent blank (water sample)	0	0	passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2021-05-07



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com