

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Glucose/D-Fructose

REF	Art. No.:	RCS4160
LOT	Lot:	2003-2
	Expiry:	2021-04

Content	Lot	Expiry
Reagent 1	2003-2	2021-04
Reagent 2	2003-2	2021-04
RFID Card	2003-2	2021-04
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1600	1499	93.7
Sample 2	246.3	234.0	95.0
Reagent blank (water sample)	0	6	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	394.0	379.0	96.2
Sample 2	49.3	47.0	95.3
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2020-03-17

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com