

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE D-Glucose/D-Fructose

REF	Art. No.:	RCS4160
LOT	Lot:	2101-1
	Expiry:	2022-03

Content	Lot	Expiry
Reagent 1	2101-1	2022-03
Reagent 2	2101-1	2022-03
RFID Card	2101-1	2022-03
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1663.8	1526	91.7
Sample 2	249.5	240	96.2
Reagent blank (water sample)	0	1	Passed

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	399.2	381	95.4
Sample 2	49.9	47	94.2
Reagent blank (water sample)	0	0	Passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-02-01

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com