


Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Sucrose/ D-Glucose/ D-Fructose

REF	Art. No.:	RCS4190
LOT	Lot:	2102-3
	Expiry:	2022-05

Content	Lot	Expiry
Reagent 1	2102-03	2022-05
Reagent 2	2102-03	2022-05
RFID Card	2102-03	2022-05
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	1482.4	1446	97.5
Sample 2	296.5	283	95.4
Reagent blank (water sample)	0	1	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	296.5	285	96.1
Sample 2	74.1	67	90.4
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-02-24



QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com