

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Acetic Acid

REF	Art. No.:	RCS4226
LOT	Lot:	2011-2
	Expiry:	2022-02

Content	Lot	Expiry
Reagent 1	2011-2	2022-02
Reagent 2	2011-2	2022-02
Reagent 3	2011-2	2022-02
RFID Card	2011-2	2022-02
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	525.0	502.0	95.6
Sample 2	145.0	126.0	86.9
Reagent blank (water sample)	0.0	1.0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	101.5	100.0	98.5
Sample 2	29.0	25.0	86.2
Reagent blank (water sample)	0.0	0.0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-01-07

Quality Management System certified
• ISO 9001
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com