

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE SO2-Total**

<b>REF</b>	Art. No.:	RCS4600
<b>LOT</b>	Lot:	2006-2
	Expiry:	2021-07

Content	Lot	Expiry
Reagent 1	2006-2	2021-07
Reagent 2	2006-2	2021-07
RFID Card	2006-2	2021-07
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	250	233	93.2
Sample 2	50	47	94.0
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	50	46	92.0
Sample 2	15	14	93.3
Reagent blank (water sample)	0	0	/

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2020-07-08

Quality Management System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com