

# Quality Assurance Certificate

r-biopharm®



**Product: RIDA® CUBE Ethanol**

<b>REF</b>	Art. No.:	RCS4340
<b>LOT</b>	Lot:	12298
	Expiry:	2019-06

Content	Lot	Expiry
Reagent 1	12298	2019-06
Reagent 2	12298	2019-06
RFID Card	12298	2019-06
IFU (and other accompanying documents where applicable)		

## Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	50	50.5	101.0
Sample 2	300	292.3	97.4
Reagent blank (water sample)	0	0	/

Sensitive application (100 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	20	20.6	103.0
Sample 2	90	93.9	104.3
Reagent blank (water sample)	0	0.3	/

R-Biopharm AG, Darmstadt, Germany, certifies that this batch has been approved by the Quality Control and conforms to specifications.

**R-Biopharm AG**  
Quality Control

Date: 2018-08-13

Quality Management System certified  
• ISO 9001  
• ISO 13485

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • [www.r-biopharm.com](http://www.r-biopharm.com)

2017-01-01