

Quality Assurance Certificate

r-biopharm®



Product: RIDA® CUBE Lactose / D-Glucose

REF	Art. No.:	RCS4130
LOT	Lot:	2102-1
	Expiry:	2022-04

Content	Lot	Expiry
Reagent 1	2102-1	2022-04
Reagent 2	2102-1	2022-04
RFID Card	2102-1	2022-04
IFU (and other accompanying documents where applicable)		

Release test

Basic application (20 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	2666.7	2609	97.8
Sample 2	821.5	818	99.6
Reagent blank (water sample)	0	0	passed

Ultra-Sensitive application (200 µl)	Target value [mg/L]	Measured value [mg/L]	Recovery [%]
Sample 1	328.6	309	94.0
Sample 2	82.2	82	99.8
Reagent blank (water sample)	0	2	passed

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Quality Control

Date: 2021-03-04



QAC-Version: 2020-04-27

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com