

# Quality Assurance Certificate

r-biopharm®



**Product:** Enzytec™ Acetic Acid

<b>REF</b>	Art. No.:	E1226
<b>LOT</b>	Lot:	200810
	Expiry:	2022-09-30

Test	Specifications	Results
<b>Part 1: Specific activity</b>	L-Malate DH $\geq$ 2700 U/ml CS $\geq$ 700 U/ml Acetyl-CoA synthetase $\geq$ 22 U/ml	corresponds corresponds corresponds
<b>Part 2: pH value</b>	Bottle #1 8.2 - 8.6	corresponds
<b>Part 3: Solubilisation time</b>	Bottle #2 $\leq$ 3 min Bottle #4 $\leq$ 3 min	corresponds corresponds
<b>Performance test</b>	Target values $\pm$ 10%	corresponds

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

## Quality Control

Date: 2021-10-05

Quality  
Management  
System certified  
• ISO 9001  
• ISO 13485

QAC-Version: 2019-04-01

Remark: This document is created electronically and therefore valid without a signature.

R-Biopharm AG • An der neuen Bergstraße 17, 64297 Darmstadt, Germany • www.r-biopharm.com